

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
 Data File : VX015198.D
 Acq On : 11 Mar 2020 14:16
 Operator : JC/SP
 Sample : PB127431ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB127431ZHE#07

Quant Time: Mar 12 06:52:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	321876	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	552562	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	548410	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	268540	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	206053	53.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.92%	
35) Dibromofluoromethane	5.49	113	173057	49.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.28%	
50) Toluene-d8	8.71	98	686207	51.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.50%	
62) 4-Bromofluorobenzene	11.13	95	249691	52.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.58%	

Target Compounds						Qvalue
16) Acetone	2.43	43	12527	8.10	ug/l	96
18) Methyl Acetate	2.76	43	4402	1.16	ug/l #	74
20) Methylene Chloride	2.85	84	12903	3.57	ug/l	94
43) Isopropyl Acetate	6.43	43	88696	10.64	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
Data File : VX015198.D
Acq On : 11 Mar 2020 14:16
Operator : JC/SP
Sample : PB127431ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB127431ZHE#07

Quant Time: Mar 12 06:52:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

