

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
 Data File : VX015200.D
 Acq On : 11 Mar 2020 15:02
 Operator : JC/SP
 Sample : PB127431ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB127431ZHE#09

Quant Time: Mar 12 06:56:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	311404	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	529323	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	533880	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	260672	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	200889	54.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.76%	
35) Dibromofluoromethane	5.49	113	166351	49.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.64%	
50) Toluene-d8	8.71	98	663088	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
62) 4-Bromofluorobenzene	11.14	95	242790	53.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.18%	
Target Compounds						
16) Acetone	2.43	43	10947	7.32	ug/l	# 88
20) Methylene Chloride	2.86	84	9387	2.68	ug/l	93
43) Isopropyl Acetate	6.43	43	85529	10.71	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
Data File : VX015200.D
Acq On : 11 Mar 2020 15:02
Operator : JC/SP
Sample : PB127431ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB127431ZHE#09

Quant Time: Mar 12 06:56:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

