

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
 Data File : VX015207.D
 Acq On : 11 Mar 2020 17:40
 Operator : JC/SP
 Sample : PB127431ZHE#16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB127431ZHE#16

Quant Time: Mar 12 07:09:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	312928	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	539913	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	541828	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	271987	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	198771	53.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.08%	
35) Dibromofluoromethane	5.49	113	168932	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
50) Toluene-d8	8.71	98	673732	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
62) 4-Bromofluorobenzene	11.13	95	247703	53.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.20%	
Target Compounds						
16) Acetone	2.43	43	11252	7.48	ug/l	95
20) Methylene Chloride	2.85	84	8872	2.52	ug/l	86
43) Isopropyl Acetate	6.43	43	83701	10.28	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
Data File : VX015207.D
Acq On : 11 Mar 2020 17:40
Operator : JC/SP
Sample : PB127431ZHE#16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB127431ZHE#16

Quant Time: Mar 12 07:09:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

