

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031121\
 Data File : VX021150.D
 Acq On : 11 Mar 2021 12:48
 Operator : JC/MD
 Sample : M1582-08
 Misc : 5.50g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VEOLIA-72-11991

Quant Time: Mar 12 02:18:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	191100	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	349209	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.100	117	333676	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	157347	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	127024	49.21	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.42%
35) Dibromofluoromethane	5.464	113	109501	51.61	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.22%
50) Toluene-d8	8.694	98	439307	53.10	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.20%
62) 4-Bromofluorobenzene	11.124	95	165767	54.19	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.38%
Target Compounds						
18) Methyl Acetate	2.758	43	5495	2.18	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031121\
Data File : VX021150.D
Acq On : 11 Mar 2021 12:48
Operator : JC/MD
Sample : M1582-08
Misc : 5.50g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VEOLIA-72-11991

Quant Time: Mar 12 02:18:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

