

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031121\
 Data File : VX021155.D
 Acq On : 11 Mar 2021 14:55
 Operator : JC/MD
 Sample : M1619-08 1000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30624

Quant Time: Mar 12 02:20:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	214744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	391931	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	398554	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	195456	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	138521	47.76	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.52%
35) Dibromofluoromethane	5.458	113	118906	49.93	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.86%
50) Toluene-d8	8.694	98	496394	53.46	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.92%
62) 4-Bromofluorobenzene	11.118	95	203659	59.32	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	118.64%
Target Compounds						
16) Acetone	2.417	43	165616	138.57	ug/l	99
52) Toluene	8.765	92	28770	4.14	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031121\
Data File : VX021155.D
Acq On : 11 Mar 2021 14:55
Operator : JC/MD
Sample : M1619-08 1000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30624

Quant Time: Mar 12 02:20:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

