

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031121\
 Data File : VX021159.D
 Acq On : 11 Mar 2021 16:26
 Operator : JC/MD
 Sample : M1647-15 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 41145

Quant Time: Mar 12 02:21:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	212910	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	387171	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	387794	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	182893	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	138561	48.18	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.36%
35) Dibromofluoromethane	5.464	113	120707	51.31	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.62%
50) Toluene-d8	8.694	98	492903	53.74	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.48%
62) 4-Bromofluorobenzene	11.118	95	190180	56.08	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	112.16%
Target Compounds						
16) Acetone	2.417	43	31695	26.75	ug/l	98
52) Toluene	8.765	92	18354	2.67	ug/l	99
68) m/p-Xylenes	10.341	106	6240	1.15	ug/l	96
69) o-Xylene	10.683	106	2811	0.53	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031121\
Data File : VX021159.D
Acq On : 11 Mar 2021 16:26
Operator : JC/MD
Sample : M1647-15 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
41145

Quant Time: Mar 12 02:21:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

