

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031121\
 Data File : VX021167.D
 Acq On : 11 Mar 2021 19:32
 Operator : JC/MD
 Sample : M1648-09 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP2-30802:05

Quant Time: Mar 12 02:24:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.634	168	209987	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.829	114	383657	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	375093	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	181939	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	137094	48.33	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.66%
35) Dibromofluoromethane	5.464	113	117807	50.54	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.08%
50) Toluene-d8	8.694	98	482187	53.05	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.10%
62) 4-Bromofluorobenzene	11.117	95	188199	56.00	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	112.00%
Target Compounds						
52) Toluene	8.764	92	3691	0.54	ug/l	95
68) m/p-Xylenes	10.341	106	4553	0.86	ug/l	97
69) o-Xylene	10.682	106	5859	1.14	ug/l	99
78) n-propylbenzene	11.341	91	19029	1.16	ug/l	99
80) 1,3,5-Trimethylbenzene	11.488	105	25647	2.14	ug/l	99
84) 1,2,4-Trimethylbenzene	11.788	105	85305	7.03	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031121\
Data File : VX021167.D
Acq On : 11 Mar 2021 19:32
Operator : JC/MD
Sample : M1648-09 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP2-30802:05

Quant Time: Mar 12 02:24:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

