


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031121\
 Data File : VX021144.D
 Acq On : 11 Mar 2021 09:47
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Title : SW846 8260
 Last Update : Wed Feb 24 13:50:55 2021



AutoFind: Scans 1704, 1705, 1706; Background Corrected with Scan 1698

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	8366	PASS
75	95	30	60	47.4	22960	PASS
95	95	100	100	100.0	48411	PASS
96	95	5	9	6.9	3320	PASS
173	174	0.00	2	0.8	295	PASS
174	95	50	100	73.3	35493	PASS
175	174	5	9	7.5	2650	PASS
176	174	95	101	98.1	34827	PASS
177	176	5	9	6.3	2211	PASS