

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031122\
 Data File : VX027404.D
 Acq On : 11 Mar 2022 16:06
 Operator : JC/MD
 Sample : N1815-02DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PW-1DL

Quant Time: Mar 11 23:12:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	384532	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	617850	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	582055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	277686	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	197586	44.062	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.120%
35) Dibromofluoromethane	5.391	113	196645	49.325	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.640%
50) Toluene-d8	8.653	98	730550	48.538	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.080%
62) 4-Bromofluorobenzene	11.085	95	272713	47.541	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.080%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.471	56	15437	2.461	ug/l	91
39) Methylcyclohexane	7.385	83	9419	1.348	ug/l	93
67) Ethyl Benzene	10.195	91	5645	0.267	ug/l	100
68) m/p-Xylenes	10.305	106	8222	0.993	ug/l	100
69) o-Xylene	10.640	106	2493	0.305	ug/l	88
80) 1,3,5-Trimethylbenzene	11.451	105	145804	8.374	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	7136	0.413	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031122\
Data File : VX027404.D
Acq On : 11 Mar 2022 16:06
Operator : JC/MD
Sample : N1815-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PW-1DL

Quant Time: Mar 11 23:12:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

