

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031122\
 Data File : VX027408.D
 Acq On : 11 Mar 2022 17:40
 Operator : JC/MD
 Sample : N1895-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB030922

Quant Time: Mar 11 23:14:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	403861	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	640463	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	595724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	279687	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	198630	42.175	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	84.360%
35) Dibromofluoromethane	5.391	113	198182	47.955	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.920%
50) Toluene-d8	8.653	98	749081	48.012	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.020%
62) 4-Bromofluorobenzene	11.085	95	275237	46.287	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.580%
Target Compounds						
					Qvalue	
52) Toluene	8.726	92	8982	0.806	ug/l	95
67) Ethyl Benzene	10.195	91	54282	2.511	ug/l	100
68) m/p-Xylenes	10.305	106	114964	13.560	ug/l	97
69) o-Xylene	10.646	106	38687	4.622	ug/l	99
78) n-propylbenzene	11.305	91	8603	0.374	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	13508	0.770	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	49558	2.849	ug/l	96
95) Naphthalene	13.780	128	27729	1.316	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031122\
Data File : VX027408.D
Acq On : 11 Mar 2022 17:40
Operator : JC/MD
Sample : N1895-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB030922

Quant Time: Mar 11 23:14:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

