

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031125\
 Data File : VX045213.D
 Acq On : 11 Mar 2025 10:51
 Operator : JC/MD
 Sample : VX0311WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0311WBL01

Quant Time: Mar 12 01:43:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	69956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	139206	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50068	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59358	53.343	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.680%
35) Dibromofluoromethane	5.385	113	47150	50.653	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.300%
50) Toluene-d8	8.647	98	170914	50.647	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.300%
62) 4-Bromofluorobenzene	11.079	95	57517	51.435	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031125\
Data File : VX045213.D
Acq On : 11 Mar 2025 10:51
Operator : JC/MD
Sample : VX0311WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0311WBL01

Quant Time: Mar 12 01:43:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

