

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031125\
 Data File : VX045223.D
 Acq On : 11 Mar 2025 14:46
 Operator : JC/MD
 Sample : Q1531-11DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D2-20250306DL

Quant Time: Mar 12 01:49:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71410	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	143079	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	131881	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	60827	53.551	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.100%	
35) Dibromofluoromethane	5.385	113	49390	51.624	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.240%	
50) Toluene-d8	8.647	98	180581	52.063	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.120%	
62) 4-Bromofluorobenzene	11.079	95	60414	52.564	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.120%	
Target Compounds						Qvalue
16) Acetone	2.379	43	401	0.761	ug/l	99
44) Trichloroethene	7.122	130	33027	33.948	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031125\
Data File : VX045223.D
Acq On : 11 Mar 2025 14:46
Operator : JC/MD
Sample : Q1531-11DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D2-20250306DL

Quant Time: Mar 12 01:49:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

