

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031125\
 Data File : VX045229.D
 Acq On : 11 Mar 2025 17:06
 Operator : JC/MD
 Sample : Q1541-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 205703

Quant Time: Mar 12 01:52:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	71218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	141227	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54624	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60769	53.644	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.280%	
35) Dibromofluoromethane	5.385	113	48874	51.754	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.500%	
50) Toluene-d8	8.646	98	174999	51.116	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.240%	
62) 4-Bromofluorobenzene	11.079	95	58710	51.751	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.500%	
Target Compounds						
16) Acetone	2.385	43	3162	6.016	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031125\
Data File : VX045229.D
Acq On : 11 Mar 2025 17:06
Operator : JC/MD
Sample : Q1541-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
205703

Quant Time: Mar 12 01:52:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

