

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031219\
 Data File : VX008010.D
 Acq On : 12 Mar 2019 12:10
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Mar 13 05:17:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	252380	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	352062	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	314135	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	160730	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	116936	43.92	ug/l	0.00
Spiked Amount			Recovery	=	87.84%	
35) Dibromofluoromethane	5.50	113	105738	46.95	ug/l	0.00
Spiked Amount			Recovery	=	93.90%	
50) Toluene-d8	8.71	98	401648	48.56	ug/l	0.00
Spiked Amount			Recovery	=	97.12%	
62) 4-Bromofluorobenzene	11.13	95	142223	49.05	ug/l	0.00
Spiked Amount			Recovery	=	98.10%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.20	85	101332	46.995	ug/l	100
3) Chloromethane	1.33	50	124477	44.796	ug/l	99
4) Vinyl Chloride	1.41	62	129057	46.743	ug/l	100
5) Bromomethane	1.64	94	67013	51.760	ug/l	100
6) Chloroethane	1.72	64	80561	43.922	ug/l	99
7) Trichlorofluoromethane	1.93	101	173355	48.225	ug/l	99
8) Diethyl Ether	2.19	74	73939	47.500	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	2.39	101	110744	49.714	ug/l	99
10) Methyl Iodide	2.51	142	160413	50.760	ug/l	99
11) Tert butyl alcohol	3.07	59	136314	228.947	ug/l	100
12) 1,1-Dichloroethene	2.37	96	103006	47.971	ug/l	97
13) Acrolein	2.29	56	116391	344.848	ug/l	98
14) Allyl chloride	2.73	41	204847	45.399	ug/l	99
15) Acrylonitrile	3.15	53	337040	224.269	ug/l	100
16) Acetone	2.45	43	358270	232.493	ug/l	99
17) Carbon Disulfide	2.57	76	304653	46.691	ug/l	100
18) Methyl Acetate	2.78	43	156765	47.885	ug/l	98
19) Methyl tert-butyl Ether	3.20	73	359143	48.617	ug/l	100
20) Methylene Chloride	2.85	84	115210	46.976	ug/l	99
21) trans-1,2-Dichloroethene	3.17	96	110951	46.960	ug/l	96
22) Diisopropyl ether	3.87	45	380670	46.024	ug/l	96
23) Vinyl Acetate	3.82	43	1636961	225.689	ug/l	98
24) 1,1-Dichloroethane	3.70	63	206541	46.402	ug/l	99
25) 2-Butanone	4.70	43	492145	225.584	ug/l	99
26) 2,2-Dichloropropane	4.58	77	152420	51.687	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	125334	46.358	ug/l	99
28) Bromochloromethane	5.02	49	93279	44.951	ug/l	97
29) Tetrahydrofuran	5.15	42	303732	223.195	ug/l	99
30) Chloroform	5.21	83	199340	47.344	ug/l	99
31) Cyclohexane	5.57	56	193364	47.267	ug/l	99
32) 1,1,1-Trichloroethane	5.49	97	169756	48.593	ug/l	98
36) 1,1-Dichloropropene	5.80	75	154794	48.558	ug/l	99
37) Ethyl Acetate	4.85	43	172285	45.542	ug/l	99
38) Carbon Tetrachloride	5.78	117	149858	49.516	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031219\
 Data File : VX008010.D
 Acq On : 12 Mar 2019 12:10
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Mar 13 05:17:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	202756	52.489	ug/l	98
40) Benzene	6.14	78	469685	49.491	ug/l	100
41) Methacrylonitrile	5.06	41	99465	48.226	ug/l	99
42) 1,2-Dichloroethane	6.20	62	150467	48.480	ug/l	99
43) Isopropyl Acetate	6.46	43	269861	48.201	ug/l	99
44) Trichloroethene	7.21	130	131240	50.180	ug/l	98
45) 1,2-Dichloropropane	7.51	63	124055	49.491	ug/l	98
46) Dibromomethane	7.66	93	77534	48.877	ug/l	98
47) Bromodichloromethane	7.90	83	148807	49.478	ug/l	99
48) Methyl methacrylate	7.77	41	142026	48.250	ug/l	98
49) 1,4-Dioxane	7.75	88	58072	960.384	ug/l	98
51) 4-Methyl-2-Pentanone	8.65	43	864513	240.106	ug/l	99
52) Toluene	8.78	92	295916	51.072	ug/l	99
53) t-1,3-Dichloropropene	9.04	75	168136	53.562	ug/l	99
54) cis-1,3-Dichloropropene	8.43	75	185859	52.431	ug/l	99
55) 1,1,2-Trichloroethane	9.21	97	119171	50.808	ug/l	99
56) Ethyl methacrylate	9.18	69	181454	51.525	ug/l	97
57) 1,3-Dichloropropane	9.37	76	196419	49.874	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.31	63	487835	267.793	ug/l	98
59) 2-Hexanone	9.49	43	664211	237.609	ug/l	99
60) Dibromochloromethane	9.58	129	125464	51.176	ug/l	99
61) 1,2-Dibromoethane	9.67	107	127287	51.113	ug/l	99
64) Tetrachloroethene	9.34	164	134248	51.570	ug/l	99
65) Chlorobenzene	10.13	112	324817	49.413	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.22	131	119698	50.807	ug/l	100
67) Ethyl Benzene	10.25	91	554527	51.247	ug/l	100
68) m/p-Xylenes	10.36	106	427797	102.569	ug/l	100
69) o-Xylene	10.69	106	205008	50.912	ug/l	100
70) Styrene	10.71	104	341166	51.776	ug/l	100
71) Bromoform	10.85	173	98539	48.704	ug/l	# 100
73) Isopropylbenzene	11.02	105	560687	51.579	ug/l	100
74) N-amyl acetate	10.90	43	228560	46.056	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.27	83	172487	46.001	ug/l	100
76) 1,2,3-Trichloropropane	11.29	75	203336	60.287	ug/l	91
77) Bromobenzene	11.26	156	147361	49.445	ug/l	98
78) n-propylbenzene	11.36	91	630015	52.242	ug/l	100
79) 2-Chlorotoluene	11.42	91	369312	50.087	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	466975	52.194	ug/l	100
81) trans-1,4-Dichloro-2-buten	11.07	75	53359	50.570	ug/l	99
82) 4-Chlorotoluene	11.51	91	426909	50.522	ug/l	100
83) tert-Butylbenzene	11.77	119	460317	51.077	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	475093	51.800	ug/l	100
85) sec-Butylbenzene	11.94	105	563707	53.109	ug/l	100
86) p-Isopropyltoluene	12.06	119	511186	53.776	ug/l	100
87) 1,3-Dichlorobenzene	12.02	146	263771	50.027	ug/l	99
88) 1,4-Dichlorobenzene	12.10	146	263149	48.735	ug/l	99
89) n-Butylbenzene	12.38	91	437010	54.264	ug/l	100
90) Hexachloroethane	12.60	117	81309	51.159	ug/l	100
91) 1,2-Dichlorobenzene	12.39	146	258795	49.408	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	13.00	75	35690	45.544	ug/l	96


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031219\
Data File : VX008010.D
Acq On : 12 Mar 2019 12:10
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Quant Time: Mar 13 05:17:26 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.65	180	193028	52.459	ug/l	100
94) Hexachlorobutadiene	13.78	225	104682	54.415	ug/l	99
95) Naphthalene	13.83	128	546711	51.566	ug/l	100
96) 1,2,3-Trichlorobenzene	14.02	180	189812	51.783	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050

Tot Ion:168 Resp: 252380

Ion Ratio Lower Upper

168 100

99 47.2 37.6 56.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

100000

80000

60000



40000

20000

0

5.50 5.60 5.70 5.80 5.90

Time-->



#2

Dichlorodifluoromethane

Concen: 46.995 ug/l

RT: 1.20 min Scan# 7

Delta R.T. 0.00 min

Lab File: VX008010.D


Acq: 12 Mar 2019 12:10

Tgt Ion: 85 Resp: 101332

Ion Ratio Lower Upper

85 100

87 32.6 16.4 49.1



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.20

100000

80000

60000


40000

20000

0

1.15 1.20 1.25 1.30

Time-->





#3

Chloromethane

Concen: 44.796 ug/l

RT: 1.33 min Scan# 29

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tot Ion: 50 Resp: 124477

Ion Ratio Lower Upper

50 100

52 32.3 26.5 39.7



#4

Vinyl Chloride

Concen: 46.743 ug/l

RT: 1.41 min Scan# 42

Delta R.T. 0.00 min

Lab File: VX008010.D




Acq: 12 Mar 2019 12:10


Tgt Ion: 62 Resp: 129057

Ion Ratio Lower Upper

62 100

64 31.4 25.3 37.9





#5

Bromomethane

Concen: 51.760 ug/l

RT: 1.64 min Scan# 80

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tot Ion: 94 Resp: 67013

Ion Ratio Lower Upper

94 100

96 94.5 75.5 113.3



#6

Chloroethane

Concen: 43.922 ug/l

RT: 1.72 min Scan# 93

Delta R.T. 0.01 min

Lab File: VX008010.D




Acq: 12 Mar 2019 12:10


Tgt Ion: 64 Resp: 80561

Ion Ratio Lower Upper

64 100

66 32.2 25.2 37.8









#7

Trichlorofluoromethane
Concen: 48.225 ug/l
RT: 1.93 min Scan# 127
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tot Ion:101 Resp: 173355
Ion Ratio Lower Upper
101 100
103 63.6 50.4 75.6




#8

Diethyl Ether
Concen: 47.500 ug/l
RT: 2.19 min Scan# 170
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 74 Resp: 73939
Ion Ratio Lower Upper
74 100
45 102.9 52.1 156.4





#9

1,1,2-Trichlorotrifluoroethane

Concen: 49.714 ug/l

RT: 2.39 min Scan# 202

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tot Ion:101 Resp: 110744

Ion Ratio Lower Upper

101 100

85 41.8 33.8 50.6




151 87.3 69.1 103.7



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 50.760 ug/l

RT: 2.51 min Scan# 223

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion:142 Resp: 160413

Ion Ratio Lower Upper

142 100

127 39.3 32.1 48.1



141 14.4 11.4 17.2




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11
Tert butyl alcohol
Concen: 228.947 ug/l
RT: 3.07 min Scan# 315
Delta R.T. -0.01 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 59 Resp: 136314
Ion Ratio Lower Upper
59 100
57 10.0 8.2 12.2



#12
1,1-Dichloroethene
Concen: 47.971 ug/l
RT: 2.37 min Scan# 200
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 96 Resp: 103006
Ion Ratio Lower Upper
96 100
61 156.3 129.8 194.6
98 64.5 51.0 76.6





#13

Acrolein

Concen: 344.848 ug/l

RT: 2.29 min Scan# 187

Delta R.T. -0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tot Ion: 56 Resp: 116391

Ion Ratio Lower Upper

56 100

55 70.6 57.8 86.6



#14

Allyl chloride

Concen: 45.399 ug/l

RT: 2.73 min Scan# 258

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10




Tgt Ion: 41 Resp: 204847

Ion Ratio Lower Upper

41 100

39 57.6 46.1 69.1

76 31.0 24.0 36.0





#15

Acrylonitrile

Concen: 224.269 ug/l

RT: 3.15 min Scan# 327

Delta R.T. -0.01 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :


MSVOA_X

ClientSampleId :

VSTDCCC050

Tot Ion: 53 Resp: 337040



Ion	Ratio	Lower	Upper
53	100		
52	83.1	66.0	99.0
51	35.0	28.0	42.0



Abundance Ion 53.00 (52.70 to 53.70): VX0

Ion 52.00 (51.70 to 52.70): VX0

Ion 51.00 (50.70 to 51.70): VX0



#16

Acetone

Concen: 232.493 ug/l

RT: 2.45 min Scan# 212


Delta R.T. -0.01 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion: 43 Resp: 358270


Ion	Ratio	Lower	Upper
43	100		
58	32.3	25.5	38.3



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0





#17

Carbon Disulfide

Concen: 46.691 ug/l

RT: 2.57 min Scan# 232

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050

Tot Ion: 76 Resp: 304653

Ion Ratio Lower Upper

76 100

78 9.0 7.2 10.8



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0

2.57

200000



150000

100000

50000

0

Time-->



#18

Methyl Acetate

Concen: 47.885 ug/l

RT: 2.78 min Scan# 266

Delta R.T. -0.01 min

Lab File: VX008010.D


Acq: 12 Mar 2019 12:10

Tgt Ion: 43 Resp: 156765

Ion Ratio Lower Upper

43 100

74 23.2 17.8 26.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0

2.78

100000

80000


60000


40000

20000

0

Time-->









#19
Methvl tert-butvl Ether
Concen: 48.617 ug/l
RT: 3.20 min Scan# 336
Delta R.T. -0.01 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 73 Resp: 359143
Ion Ratio Lower Upper
73 100
57 23.0 18.2 27.4



#20
Methylene Chloride
Concen: 46.976 ug/l
RT: 2.85 min Scan# 279
Delta R.T. -0.01 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 84 Resp: 115210
Ion Ratio Lower Upper
84 100
49 125.8 99.3 148.9
51 38.3 31.3 46.9
86 65.4 52.6 79.0





#21

trans-1,2-Dichloroethene

Concen: 46.960 ug/l

RT: 3.17 min Scan# 330

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tot Ion: 96 Resp: 110951

Ion Ratio Lower Upper

96 100

61 135.8 114.4 171.6

98 63.3 51.1 76.7






Abundance

Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 46.024 ug/l

RT: 3.87 min Scan# 445

Delta R.T. -0.01 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Tgt Ion: 45 Resp: 380670


Ion Ratio Lower Upper

45 100

43 61.0 45.3 67.9

87 26.7 21.4 32.0

59 11.7 9.5 14.3





Abundance


Ion 45.00 (44.70 to 45.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 87.00 (86.70 to 87.70): VX0

Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 225.689 ug/l

RT: 3.82 min Scan# 438

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tot Ion: 43 Resp: 1636961

Ion Ratio Lower Upper

43 100

86 10.5 7.9 11.9



#24

1,1-Dichloroethane

Concen: 46.402 ug/l

RT: 3.70 min Scan# 417

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10




Tgt Ion: 63 Resp: 206541


Ion Ratio Lower Upper

63 100

98 6.8 3.3 9.8

100 4.4 2.3 6.8





#25

2-Butanone

Concen: 225.584 ug/l

RT: 4.70 min Scan# 581

Delta R.T. -0.01 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 492145

Ion Ratio Lower Upper

43 100

72 25.6 19.9 29.9



Abundance Ion 43.00 (42.70 to 43.70): VX007854.D (-569) (-)

Ion 72.00 (71.70 to 72.70): VX007854.D (-569) (-)

200000

150000



100000

50000

0

4.60 4.70 4.80

Time-->



#26

2,2-Dichloropropane

Concen: 51.687 ug/l

RT: 4.58 min Scan# 562

Delta R.T. -0.01 min

Lab File: VX008010.D


Acq: 12 Mar 2019 12:10

Tgt Ion: 77 Resp: 152420

Ion Ratio Lower Upper

77 100

97 24.7 12.2 36.6



Abundance Ion 76.90 (76.60 to 77.60): VX007854.D (-549) (-)

Ion 96.90 (96.60 to 97.60): VX007854.D (-549) (-)

50000

40000

30000


20000

10000

0

4.40 4.50 4.60 4.70

Time-->






#27




cis-1,2-Dichloroethene
Concen: 46.358 ug/l
RT: 4.60 min Scan# 565
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 96 Resp: 125334
Ion Ratio Lower Upper
96 100
61 142.6 0.0 288.0
98 64.8 0.0 129.0




Abundance Ion 95.90 (95.60 to 96.60): VX0
Ion 60.90 (60.60 to 61.60): VX0
Ion 97.90 (97.60 to 98.60): VX0





#28


Bromochloromethane
Concen: 44.951 ug/l
RT: 5.02 min Scan# 634
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 49 Resp: 93279
Ion Ratio Lower Upper
49 100
129 2.1 0.0 4.0
130 84.1 65.0 97.4



Abundance Ion 48.90 (48.60 to 49.60): VX0
Ion 128.80 (128.50 to 129.50): V
Ion 129.80 (129.50 to 130.50): V









#29
Tetrahydrofuran
Concen: 223.195 ug/l
RT: 5.15 min Scan# 655
Delta R.T. -0.01 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tot Ion: 42 Resp: 303732
Ion Ratio Lower Upper
42 100
72 43.0 33.8 50.6
71 40.2 31.5 47.3



#30
Chloroform
Concen: 47.344 ug/l
RT: 5.21 min Scan# 666
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 83 Resp: 199340
Ion Ratio Lower Upper
83 100
85 65.5 51.8 77.6





#31

Cyclohexane

Concen: 47.267 ug/l

RT: 5.57 min Scan# 725

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tot Ion: 56 Resp: 193364

Ion Ratio Lower Upper




56 100

69 29.5 24.0 36.0

84 82.8 65.3 97.9



Abundance Ion 56.00 (55.70 to 56.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 84.00 (83.70 to 84.70): VX0



#32

1,1,1-Trichloroethane

Concen: 48.593 ug/l

RT: 5.49 min Scan# 712

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion: 97 Resp: 169756

Ion Ratio Lower Upper



97 100


99 64.1 52.2 78.2

61 41.3 34.9 52.3



Abundance Ion 96.90 (96.60 to 97.60): VX0
Ion 98.90 (98.60 to 99.60): VX0
Ion 60.90 (60.60 to 61.60): VX0





#33

1,2-Dichloroethane-d4

Concen: 43.922 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050

Tot Ion: 65 Resp: 116936

Ion Ratio Lower Upper

65 100

67 55.0 0.0 110.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

50000

40000

30000



20000

10000

0

6.00 6.10 6.20

Time-->



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion: 114 Resp: 352062

Ion Ratio Lower Upper

114 100

63 19.1 0.0 39.8

88 15.6 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0

6.86

200000

150000


100000


50000

0

6.70 6.80 6.90 7.00

Time-->





#35

Dibromofluoromethane

Concen: 46.952 ug/l

RT: 5.50 min Scan# 713

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tot Ion:113 Resp: 105738

Ion Ratio Lower Upper

113 100

111 101.3 81.4 122.0




192 24.4 19.4 29.0



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 48.558 ug/l

RT: 5.80 min Scan# 762

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion: 75 Resp: 154794

Ion Ratio Lower Upper

75 100

110 37.3 18.4 55.0



77 31.0 25.0 37.4




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0









#37
Ethyl Acetate
Concen: 45.542 ug/l
RT: 4.85 min Scan# 607
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 43 Resp: 172285
Ion Ratio Lower Upper
43 100
61 14.4 11.3 16.9
70 10.8 8.5 12.7




Abundance Ion 43.00 (42.70 to 43.70): VX0
Ion 60.90 (60.60 to 61.60): VX0
Ion 70.00 (69.70 to 70.70): VX0





#38
Carbon Tetrachloride
Concen: 49.516 ug/l
RT: 5.78 min Scan# 759
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 117 Resp: 149858
Ion Ratio Lower Upper
117 100
119 97.7 76.2 114.2
121 29.9 24.6 36.8



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V









#39
Methvlcvclohexane
Concen: 52.489 ug/l
RT: 7.46 min Scan# 1034
Delta R.T. -0.01 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tot Ion: 83 Resp: 202756
Ion Ratio Lower Upper
83 100
55 77.3 60.4 90.6
98 44.7 36.5 54.7



#40
Benzene
Concen: 49.491 ug/l
RT: 6.14 min Scan# 818
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 78 Resp: 469685
Ion Ratio Lower Upper
78 100
77 23.6 19.0 28.4









#41
Methacrylonitrile
Concen: 48.226 ug/l
RT: 5.06 min Scan# 640
Delta R.T. -0.01 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tot Ion:	41	Resp:	99465
Ion	Ratio	Lower	Upper
41	100		
39	51.3	41.6	62.4
67	70.3	55.8	83.6
52	26.9	21.7	32.5



#42
1,2-Dichloroethane
Concen: 48.480 ug/l
RT: 6.20 min Scan# 827
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion:	62	Resp:	150467
Ion	Ratio	Lower	Upper
62	100		
98	10.5	0.0	20.4






#43




Isopropyl Acetate
 Concen: 48.201 ug/l
 RT: 6.46 min Scan# 870
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion	43	Resp	269861
Ion Ratio	Lower	Upper	
43	100		
61	21.0	16.3	24.5
87	13.4	10.5	15.7




Abundance Ion 43.00 (42.70 to 43.70): VX0
 Ion 61.00 (60.70 to 61.70): VX0
 Ion 87.00 (86.70 to 87.70): VX0





#44


Trichloroethene
 Concen: 50.180 ug/l
 RT: 7.21 min Scan# 994
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10

Tgt Ion	130	Resp	131240
Ion Ratio <td>Lower</td> <td>Upper</td> <td></td>	Lower	Upper	
130	100		
95	92.0	0.0	188.0



Abundance Ion 129.80 (129.50 to 130.50): V
 Ion 94.90 (94.60 to 95.60): VX0









#45
 1,2-Dichloropropane
 Concen: 49.491 ug/l
 RT: 7.51 min Scan# 1043
 Delta R.T. -0.01 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tot Ion: 63 Resp: 124055
 Ion Ratio Lower Upper
 63 100
 65 29.9 24.7 37.1



#46
 Dibromomethane
 Concen: 48.877 ug/l
 RT: 7.66 min Scan# 1067
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10

Tgt Ion: 93 Resp: 77534
 Ion Ratio Lower Upper
 93 100
 95 84.2 66.8 100.2
 174 125.7 97.3 145.9





#47

Bromodichloromethane

Concen: 49.478 ug/l

RT: 7.90 min Scan# 1106

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tot Ion: 83 Resp: 148807

Ion Ratio Lower Upper

83 100

85 64.5 50.9 76.3




127 9.0 7.4 11.0



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 48.250 ug/l

RT: 7.77 min Scan# 1086

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion: 41 Resp: 142026

Ion Ratio Lower Upper

41 100

69 84.3 65.2 97.8



39 48.9 39.4 59.2




Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0









#49
1,4-Dioxane
Concen: 960.384 ug/l
RT: 7.75 min Scan# 1082
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 88 Resp: 58072
Ion Ratio Lower Upper
88 100
43 30.5 25.1 37.7
58 70.5 57.8 86.6



#50
Toluene-d8
Concen: 48.563 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 98 Resp: 401648
Ion Ratio Lower Upper
98 100
100 64.9 51.7 77.5





#51

4-Methyl-2-Pentanone

Concen: 240.106 ug/l

RT: 8.65 min Scan# 1229

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 864513

Ion Ratio Lower Upper




43 100

58 39.7 31.3 46.9



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#52

Toluene

Concen: 51.072 ug/l

RT: 8.78 min Scan# 1251

Delta R.T. -0.01 min

Lab File: VX008010.D


Acq: 12 Mar 2019 12:10

Tgt Ion: 92 Resp: 295916

Ion Ratio Lower Upper



92 100


91 170.8 137.4 206.0



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0





#53

t-1,3-Dichloropropene

Concen: 53.562 ug/l

RT: 9.04 min Scan# 1294

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tot Ion: 75 Resp: 168136

Ion Ratio Lower Upper

75 100

77 32.2 25.4 38.0



#54

cis-1,3-Dichloropropene

Concen: 52.431 ug/l

RT: 8.43 min Scan# 1194

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10




Tgt Ion: 75 Resp: 185859


Ion Ratio Lower Upper

75 100

77 31.1 25.0 37.6

39 44.5 36.6 54.8









#55
 1,1,2-Trichloroethane
 Concen: 50.808 ug/l
 RT: 9.21 min Scan# 1322
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Tot Ion:	97	Resp:	119171
Ion	Ratio	Lower	Upper
97	100		
83	86.2	69.1	103.7
85	55.3	44.3	66.5
99	64.0	49.8	74.8




Abundance Ion 96.90 (96.60 to 97.60): VX0
 Ion 82.90 (82.60 to 83.60): VX0
 Ion 84.90 (84.60 to 85.60): VX0
 Ion 98.90 (98.60 to 99.60): VX0






#56
 Ethyl methacrylate
 Concen: 51.525 ug/l
 RT: 9.18 min Scan# 1316
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10

Tgt Ion:	69	Resp:	181454
Ion	Ratio	Lower	Upper
69	100		
41	65.6	54.2	81.4
39	29.9	25.0	37.4



Abundance Ion 69.00 (68.70 to 69.70): VX0
 Ion 41.00 (40.70 to 41.70): VX0
 Ion 39.00 (38.70 to 39.70): VX0





#57

1,3-Dichloropropane

Concen: 49.874 ug/l

RT: 9.37 min Scan# 1348

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050

Tot Ion: 76 Resp: 196419

Ion Ratio Lower Upper




76 100

78 32.0 25.5 38.3



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0



#58

2-Chloroethyl Vinyl ether

Concen: 267.793 ug/l

RT: 8.31 min Scan# 1174

Delta R.T. 0.00 min

Lab File: VX008010.D


Acq: 12 Mar 2019 12:10

Tgt Ion: 63 Resp: 487835

Ion Ratio Lower Upper



63 100


106 28.5 22.2 33.2



Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 105.90 (105.60 to 106.60): V









#59
2-Hexanone
Concen: 237.609 ug/l
RT: 9.49 min Scan# 1368
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tot Ion: 43 Resp: 664211
Ion Ratio Lower Upper
43 100
58 56.0 27.6 82.7



#60
Dibromochloromethane
Concen: 51.176 ug/l
RT: 9.58 min Scan# 1382
Delta R.T. -0.01 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 129 Resp: 125464
Ion Ratio Lower Upper
129 100
127 77.4 38.3 114.8





#61

1,2-Dibromoethane

Concen: 51.113 ug/l

RT: 9.67 min Scan# 1397

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050

Tot Ion:107 Resp: 127287

Ion Ratio Lower Upper

107 100

109 94.1 75.8 113.8



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V

9.67

100000

80000



60000

40000

20000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 49.047 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion: 95 Resp: 142223

Ion Ratio Lower Upper

95 100

174 93.4 0.0 182.8

176 90.4 0.0 178.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

11.13


150000


100000

50000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tot Ion:117 Resp: 314135
Ion Ratio Lower Upper
117 100
82 54.0 44.3 66.5
119 32.3 25.3 37.9



#64
Tetrachloroethene
Concen: 51.570 ug/l
RT: 9.34 min Scan# 1342
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion:164 Resp: 134248
Ion Ratio Lower Upper
164 100
166 128.8 103.4 155.0
129 89.4 72.2 108.4
131 84.7 69.4 104.2









#65
Chlorobenzene
Concen: 49.413 ug/l
RT: 10.13 min Scan# 1473
Delta R.T. -0.01 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tot Ion:112 Resp: 324817
Ion Ratio Lower Upper
112 100
114 31.8 26.0 39.0



#66
1,1,1,2-Tetrachloroethane
Concen: 50.807 ug/l
RT: 10.22 min Scan# 1487
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion:131 Resp: 119698
Ion Ratio Lower Upper
131 100
133 95.6 47.8 143.3
119 64.9 32.5 97.5









#67
Ethyl Benzene
Concen: 51.247 ug/l
RT: 10.25 min Scan# 1492
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tot Ion: 91 Resp: 554527
Ion Ratio Lower Upper
91 100
106 31.5 25.2 37.8



#68
m/p-Xylenes
Concen: 102.569 ug/l
RT: 10.36 min Scan# 1510
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 106 Resp: 427797
Ion Ratio Lower Upper
106 100
91 200.2 159.8 239.6









#69
o-Xylene
Concen: 50.912 ug/l
RT: 10.69 min Scan# 1565
Delta R.T. -0.01 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion:106 Resp: 205008
Ion Ratio Lower Upper
106 100
91 212.5 106.3 318.9




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VX0






#70
Styrene
Concen: 51.776 ug/l
RT: 10.71 min Scan# 1568
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion:104 Resp: 341166
Ion Ratio Lower Upper
104 100
78 48.0 38.5 57.7
103 55.4 44.1 66.1



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VX0
Ion 103.00 (102.70 to 103.70): V





#71

Bromoform

Concen: 48.704 ug/l

RT: 10.85 min Scan# 1591

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tot Ion:173 Resp: 98539

Ion Ratio Lower Upper




173 100

175 48.7 24.3 72.9

254 0.2 0.1 0.1#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion:152 Resp: 160730

Ion Ratio Lower Upper



152 100

115 77.6 38.6 115.7

150 173.5 0.0 348.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 51.579 ug/l

RT: 11.02 min Scan# 1618

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050

Tot Ion:105 Resp: 560687

Ion Ratio Lower Upper




105 100

120 26.8 13.4 40.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 46.056 ug/l

RT: 10.90 min Scan# 1598

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Tgt Ion: 43 Resp: 228560


Ion Ratio Lower Upper

43 100

70 43.2 33.5 50.3

55 25.0 19.2 28.8

61 24.8 19.0 28.6






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0





#75

1,1,2,2-Tetrachloroethane

Concen: 46.001 ug/l

RT: 11.27 min Scan# 1659

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tot Ion: 83 Resp: 172487

Ion Ratio Lower Upper

83 100

131 10.7 5.4 16.2




85 64.3 32.3 96.8



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VX0



#76

1,2,3-Trichloropropane

Concen: 60.287 ug/l

RT: 11.29 min Scan# 1663

Delta R.T. -0.01 min

Lab File: VX008010.D


Acq: 12 Mar 2019 12:10

Tgt Ion: 75 Resp: 203336

Ion Ratio Lower Upper



75 100


77 32.5 18.9 56.7



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#77

Bromobenzene

Concen: 49.445 ug/l

RT: 11.26 min Scan# 1657

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tot Ion:156 Resp: 147361

Ion Ratio Lower Upper

156 100

77 141.0 72.0 215.9




158 97.6 48.6 145.9



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 52.242 ug/l

RT: 11.36 min Scan# 1674

Delta R.T. 0.00 min

Lab File: VX008010.D


Acq: 12 Mar 2019 12:10

Tgt Ion: 91 Resp: 630015

Ion Ratio Lower Upper



91 100

120 23.9 12.0 36.0



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V









#79
 2-Chlorotoluene
 Concen: 50.087 ug/l
 RT: 11.42 min Scan# 1684
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tot Ion: 91 Resp: 369312
 Ion Ratio Lower Upper
 91 100
 126 35.3 17.5 52.6



#80
 1,3,5-Trimethylbenzene
 Concen: 52.194 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10

Tgt Ion: 105 Resp: 466975
 Ion Ratio Lower Upper
 105 100
 120 49.6 24.9 74.6





#81

trans-1,4-Dichloro-2-butene

Concen: 50.570 ug/l

RT: 11.07 min Scan# 1627

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tot Ion: 75 Resp: 53359

Ion Ratio Lower Upper

75 100

53 93.8 76.1 114.1




89 45.4 36.3 54.5



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#82

4-Chlorotoluene

Concen: 50.522 ug/l

RT: 11.51 min Scan# 1699

Delta R.T. 0.00 min

Lab File: VX008010.D


Acq: 12 Mar 2019 12:10

Tgt Ion: 91 Resp: 426909

Ion Ratio Lower Upper



91 100


126 31.4 15.6 46.8



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V









#83
tert-Butylbenzene
Concen: 51.077 ug/l
RT: 11.77 min Scan# 1741
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tot Ion:119 Resp: 460317
Ion Ratio Lower Upper
119 100
91 55.5 27.1 81.3
134 23.2 11.4 34.1



#84
1,2,4-Trimethylbenzene
Concen: 51.800 ug/l
RT: 11.80 min Scan# 1747
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion:105 Resp: 475093
Ion Ratio Lower Upper
105 100
120 45.1 22.6 67.7





#85

sec-Butylbenzene

Concen: 53.109 ug/l

RT: 11.94 min Scan# 1770

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050

Tot Ion:105 Resp: 563707

Ion Ratio Lower Upper




105 100

134 20.6 10.4 31.2



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 53.776 ug/l

RT: 12.06 min Scan# 1789

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion:119 Resp: 511186

Ion Ratio Lower Upper

119 100

134 26.4 13.3 39.9



91 21.7 10.9 32.7




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V









#87
 1,3-Dichlorobenzene
 Concen: 50.027 ug/l
 RT: 12.02 min Scan# 1783
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tot Ion:146 Resp: 263771
 Ion Ratio Lower Upper
 146 100
 111 37.6 18.9 56.7
 148 64.3 31.7 95.1



#88
 1,4-Dichlorobenzene
 Concen: 48.735 ug/l
 RT: 12.10 min Scan# 1795
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10

Tgt Ion:146 Resp: 263149
 Ion Ratio Lower Upper
 146 100
 111 37.1 18.4 55.0
 148 64.8 32.2 96.6









#89
n-Butylbenzene
Concen: 54.264 ug/l
RT: 12.38 min Scan# 1842
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tot Ion: 91 Resp: 437010
Ion Ratio Lower Upper
91 100
92 53.6 27.0 80.8
134 27.6 13.7 41.0



#90
Hexachloroethane
Concen: 51.159 ug/l
RT: 12.60 min Scan# 1877
Delta R.T. 0.00 min
Lab File: VX008010.D
Acq: 12 Mar 2019 12:10

Tgt Ion: 117 Resp: 81309
Ion Ratio Lower Upper
117 100
201 97.9 48.9 146.7





#91

1,2-Dichlorobenzene

Concen: 49.408 ug/l

RT: 12.39 min Scan# 1843

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampled :

VSTDCCC050


Tot Ion:146 Resp: 258795

Ion Ratio Lower Upper

146 100

111 39.1 19.6 58.8




148 64.3 31.9 95.5



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 45.544 ug/l

RT: 13.00 min Scan# 1943

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion: 75 Resp: 35690

Ion Ratio Lower Upper

75 100

155 102.8 49.7 149.1



157 132.0 63.8 191.4




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93
 1,2,4-Trichlorobenzene
 Concen: 52.459 ug/l
 RT: 13.65 min Scan# 2049
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tot Ion:180 Resp: 193028
 Ion Ratio Lower Upper
 180 100
 182 95.5 47.9 143.7
 145 28.5 14.1 42.4



#94
 Hexachlorobutadiene
 Concen: 54.415 ug/l
 RT: 13.78 min Scan# 2071
 Delta R.T. 0.00 min
 Lab File: VX008010.D
 Acq: 12 Mar 2019 12:10

Tgt Ion:225 Resp: 104682
 Ion Ratio Lower Upper
 225 100
 223 63.0 31.1 93.2
 227 65.2 32.0 96.0





#95

Naphthalene

Concen: 51.566 ug/l

RT: 13.83 min Scan# 2080

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050


Tot Ion:128 Resp: 546711

Ion Ratio Lower Upper

128 100

127 12.7 10.3 15.5




129 10.8 8.7 13.1



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 51.783 ug/l

RT: 14.02 min Scan# 2110

Delta R.T. 0.00 min

Lab File: VX008010.D

Acq: 12 Mar 2019 12:10


Tgt Ion:180 Resp: 189812

Ion Ratio Lower Upper

180 100

182 95.9 47.5 142.5

145 30.0 14.8 44.3



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

