

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
 Data File : VX008016.D
 Acq On : 12 Mar 2019 15:14
 Operator : JC/SP
 Sample : K1841-01DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D2-20190308DL

Quant Time: Mar 13 10:21:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	352541	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	638735	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	628015	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	320084	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	238452	64.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.24%	
35) Dibromofluoromethane	5.50	113	232646	56.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.88%	
50) Toluene-d8	8.71	98	858191	57.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.38%	
62) 4-Bromofluorobenzene	11.13	95	321197	61.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.10%	
Target Compounds						
44) Trichloroethene	7.21	130	88173	18.582	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
Data File : VX008016.D
Acq On : 12 Mar 2019 15:14
Operator : JC/SP
Sample : K1841-01DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D2-20190308DL

Quant Time: Mar 13 10:21:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

