

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
 Data File : VX008019.D
 Acq On : 12 Mar 2019 16:34
 Operator : JC/SP
 Sample : K1841-07DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D2-20190308DL

Quant Time: Mar 13 10:23:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	232243	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	363868	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	339393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	175958	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	128052	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
35) Dibromofluoromethane	5.50	113	117928	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
50) Toluene-d8	8.71	98	440305	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
62) 4-Bromofluorobenzene	11.13	95	161901	54.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.04%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	6006	2.930	ug/l	96
12) 1,1-Dichloroethene	2.37	96	1978	1.001	ug/l	95
27) cis-1,2-Dichloroethene	4.59	96	2154	0.866	ug/l	93
44) Trichloroethene	7.21	130	143985	53.267	ug/l	98
64) Tetrachloroethene	9.34	164	2472	0.879	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
Data File : VX008019.D
Acq On : 12 Mar 2019 16:34
Operator : JC/SP
Sample : K1841-07DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D2-20190308DL

Quant Time: Mar 13 10:23:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

