

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
 Data File : VX008020.D
 Acq On : 12 Mar 2019 17:00
 Operator : JC/SP
 Sample : K1841-08DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D3-20190308DL

Quant Time: Mar 13 10:24:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	231750	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	359285	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	337880	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	174934	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	126338	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
35) Dibromofluoromethane	5.50	113	116348	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
50) Toluene-d8	8.71	98	436693	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
62) 4-Bromofluorobenzene	11.13	95	161383	54.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.08%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	15277	7.468	ug/l	100
27) cis-1,2-Dichloroethene	4.59	96	1058	0.426	ug/l	80
44) Trichloroethene	7.21	130	98542	36.920	ug/l	99
64) Tetrachloroethene	9.34	164	2019	0.721	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
Data File : VX008020.D
Acq On : 12 Mar 2019 17:00
Operator : JC/SP
Sample : K1841-08DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D3-20190308DL

Quant Time: Mar 13 10:24:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

