

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
 Data File : VX008023.D
 Acq On : 12 Mar 2019 18:20
 Operator : JC/SP
 Sample : K1841-11DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20190308DL

Quant Time: Mar 13 10:25:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252790	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	393460	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	377716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	182770	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	138896	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
35) Dibromofluoromethane	5.50	113	127821	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
50) Toluene-d8	8.71	98	486590	52.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.28%	
62) 4-Bromofluorobenzene	11.13	95	176615	54.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.00%	
Target Compounds						
44) Trichloroethene	7.21	130	184125	62.994	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
Data File : VX008023.D
Acq On : 12 Mar 2019 18:20
Operator : JC/SP
Sample : K1841-11DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20190308DL

Quant Time: Mar 13 10:25:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

