

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
 Data File : VX008026.D
 Acq On : 12 Mar 2019 19:40
 Operator : JC/SP
 Sample : K1793-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20190306

Quant Time: Mar 13 06:38:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252047	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	393768	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	373994	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	185833	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	138982	52.27	ug/l	0.00
Spiked Amount			Recovery	=	104.54%	
35) Dibromofluoromethane	5.50	113	128187	50.89	ug/l	0.00
Spiked Amount			Recovery	=	101.78%	
50) Toluene-d8	8.71	98	477766	51.65	ug/l	0.00
Spiked Amount			Recovery	=	103.30%	
62) 4-Bromofluorobenzene	11.13	95	178057	54.90	ug/l	0.00
Spiked Amount			Recovery	=	109.80%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	13975	6.282	ug/l	98
12) 1,1-Dichloroethene	2.38	96	13528	6.308	ug/l	99
16) Acetone	2.45	43	3788	2.461	ug/l	96
24) 1,1-Dichloroethane	3.70	63	16310	3.669	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	21013	7.782	ug/l	89
29) Tetrahydrofuran	5.15	42	1857	1.366	ug/l	94
30) Chloroform	5.21	83	11627	2.765	ug/l	92
32) 1,1,1-Trichloroethane	5.50	97	2209	0.633	ug/l #	48
38) Carbon Tetrachloride	5.79	117	4827	1.426	ug/l	94
44) Trichloroethene	7.23	130	9216300	3150.655	ug/l	99
55) 1,1,2-Trichloroethane	9.21	97	3866	1.474	ug/l	92
64) Tetrachloroethene	9.34	164	10543	3.402	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
Data File : VX008026.D
Acq On : 12 Mar 2019 19:40
Operator : JC/SP
Sample : K1793-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20190306

Quant Time: Mar 13 06:38:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

