

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
 Data File : VX008031.D
 Acq On : 12 Mar 2019 21:53
 Operator : JC/SP
 Sample : K1840-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-700-702

Quant Time: Mar 13 06:49:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	239213	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	374578	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	358108	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	175298	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	132265	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
35) Dibromofluoromethane	5.51	113	121891	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
50) Toluene-d8	8.71	98	458617	52.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.24%	
62) 4-Bromofluorobenzene	11.13	95	169751	55.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.04%	
Target Compounds						
16) Acetone	2.45	43	10075	6.898	ug/l	91
25) 2-Butanone	4.70	43	3757	1.817	ug/l #	90
44) Trichloroethene	7.21	130	2765	0.994	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031219\
Data File : VX008031.D
Acq On : 12 Mar 2019 21:53
Operator : JC/SP
Sample : K1840-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-700-702

Quant Time: Mar 13 06:49:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

