

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031220\
 Data File : VX015225.D
 Acq On : 12 Mar 2020 10:24
 Operator : JC/SP
 Sample : L1853-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-EC-52-G

Quant Time: Mar 13 02:30:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	289822	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	456102	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	411252	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	208714	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	163203	47.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.94%	
35) Dibromofluoromethane	5.49	113	140749	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
50) Toluene-d8	8.71	98	524879	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
62) 4-Bromofluorobenzene	11.13	95	191879	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
Target Compounds						
16) Acetone	2.43	43	23139	16.617	ug/l	96
20) Methylene Chloride	2.85	84	21906	6.722	ug/l	98
43) Isopropyl Acetate	6.43	43	39212	5.699	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031220\
Data File : VX015225.D
Acq On : 12 Mar 2020 10:24
Operator : JC/SP
Sample : L1853-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-EC-52-G

Quant Time: Mar 13 02:30:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

