

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031221\
 Data File : VX021184.D
 Acq On : 12 Mar 2021 16:33
 Operator : JC/MD
 Sample : M1677-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1


Instrument :
 MSVOA_X
 ClientSampleId :
 D2470-D2800(WATER)

Quant Time: Mar 13 01:07:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X031221W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 12 12:55:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.982	128	15169	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.829	114	84205	30.00	ug/l	# 0.00
57) Chlorobenzene-d5	10.094	117	73535	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.029	65	42526	30.46	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.53%
60) 4-Bromofluorobenzene	11.118	95	37976	28.25	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.17%
63) Toluene-d8	8.694	98	102854	30.83	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.77%
Target Compounds						
					Qvalue	
5) Bromomethane	1.623	94	346	0.56	ug/l	# 41
13) Acrolein	2.270	56	165	0.92	ug/l	# 62
15) Acetone	2.423	58	6133	42.93	ug/l	# 50
16) Carbon Disulfide	2.552	76	2033	0.82	ug/l	97
17) Allyl chloride	2.693	41	632	0.35	ug/l	# 44
18) Methylene Chloride	2.834	84	489	0.48	ug/l	87
23) tert-Butyl Alcohol	3.011	59	788	3.32	ug/l	# 99
25) Chloroform	5.176	83	42414	21.95	ug/l	98
30) 2-Butanone	4.635	43	716666	913.55	ug/l	91
38) Methylcyclohexane	7.870	83	1106	0.70	ug/l	# 18
40) Dibromomethane	7.635	93	168	0.22	ug/l	77
41) Bromodichloromethane	7.870	83	1106	0.68	ug/l	# 94
51) Ethyl methacrylate	8.700	69	444	0.27	ug/l	# 63
56) Bromoform	11.124	173	177	0.21	ug/l	# 1
58) 4-Methyl-2-Pentanone	8.618	43	798	0.54	ug/l	# 81
59) 2-Hexanone	9.471	43	265	0.24	ug/l	# 45
61) Tetrachloroethene	9.324	164	195	0.20	ug/l	# 66
62) Toluene	8.765	91	6066	1.49	ug/l	95
66) Ethyl Benzene	10.235	91	228641	48.77	ug/l	98
67) m/p-Xylenes	10.341	106	346147	205.11	ug/l	92
68) o-Xylene	10.683	106	195600	119.72	ug/l	91
69) Styrene	10.683	104	10264	3.70	ug/l	# 1
70) Isopropylbenzene	11.000	105	2012	0.45	ug/l	99
77) t-1,4-Dichloro-2-butene	11.118	75	21229	44.82	ug/l	# 11
79) tert-butylbenzene	12.047	119	1031	0.29	ug/l	73
80) 1,2,4-Trimethylbenzene	11.794	105	1060	0.28	ug/l	83
82) p-Isopropyltoluene	12.047	119	1031	0.27	ug/l	90
83) 1,3-Dichlorobenzene	12.006	146	606	0.32	ug/l	92
84) 1,4-Dichlorobenzene	12.083	146	827	0.43	ug/l	94
87) 1,2-Dichlorobenzene	12.083	146	827	0.45	ug/l	91
89) 1,2,4-Trichlorobenzene	13.624	180	590	0.50	ug/l	89
91) Naphthalene	13.818	128	6719	1.78	ug/l	99
92) 1,2,3-Trichlorobenzene	13.624	180	590	0.52	ug/l	90





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Bromochloromethane
Concen: 30.00 ug/l
RT: 4.982 min Scan# 662
Delta R.T. -0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion:128 Resp: 15169
Ion Ratio Lower Upper
128 100
49 189.3 0.0 367.5
130 130.0 0.0 318.8



#5
Bromomethane
Concen: 0.56 ug/l
RT: 1.623 min Scan# 91
Delta R.T. -0.005 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33


Tgt Ion: 94 Resp: 346
Ion Ratio Lower Upper
94 100
96 38.6 76.6 114.8#








#13
Acrolein
Concen: 0.92 ug/l
RT: 2.270 min Scan# 201
Delta R.T. -0.005 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)







Tgt Ion: 56 Resp: 165
Ion Ratio Lower Upper
56 100
55 39.4 56.3 84.5#



#15
Acetone
Concen: 42.93 ug/l
RT: 2.423 min Scan# 227
Delta R.T. 0.006 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33


Tgt Ion: 58 Resp: 6133
Ion Ratio Lower Upper
58 100
43 387.7 232.6 348.8#








#16
Carbon Disulfide
Concen: 0.82 ug/l
RT: 2.552 min Scan# 249
Delta R.T. 0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33


Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)





Tgt Ion: 76 Resp: 2033
Ion Ratio Lower Upper
76 100
78 10.3 7.4 11.0




#17
Allyl chloride
Concen: 0.35 ug/l
RT: 2.693 min Scan# 273
Delta R.T. -0.012 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33



Tgt Ion: 41 Resp: 632
Ion Ratio Lower Upper
41 100
39 29.8 34.8 104.4#
76 0.0 21.3 63.9#





#18

Methylene Chloride

Concen: 0.48 ug/l

RT: 2.834 min Scan# 297

Delta R.T. 0.000 min

Lab File: VX021184.D

Acq: 12 Mar 2021 16:33

Tgt Ion: 84 Resp: 489





Ion Ratio Lower Upper

84 100

49 130.0 90.6 136.0

51 28.2 26.3 39.5

86 55.2 51.3 76.9



#23

tert-Butyl Alcohol

Concen: 3.32 ug/l

RT: 3.011 min Scan# 327

Delta R.T. 0.006 min

Lab File: VX021184.D




Acq: 12 Mar 2021 16:33


Tgt Ion: 59 Resp: 788

Ion Ratio Lower Upper

59 100

52 2.3 0.1 0.1#









#25
Chloroform
Concen: 21.95 ug/l
RT: 5.176 min Scan# 695
Delta R.T. -0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion: 83 Resp: 42414
Ion Ratio Lower Upper
83 100
85 62.7 51.2 76.8



#27
1,2-Dichloroethane-d4
Concen: 30.46 ug/l
RT: 6.029 min Scan# 840
Delta R.T. -0.006 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Tgt Ion: 65 Resp: 42526
Ion Ratio Lower Upper
65 100
67 50.6 41.9 62.9





#28

1,4-Difluorobenzene

Concen: 30.00 ug/l

RT: 6.829 min Scan# 976





Delta R.T. 0.000 min

Lab File: VX021184.D

Acq: 12 Mar 2021 16:33

Tgt Ion:	114	Resp:	84205
Ion Ratio	Lower	Upper	
114	100		
63	24.4	16.2	24.2#
88	17.0	13.0	19.6

Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)



#30

2-Butanone

Concen: 913.55 ug/l




RT: 4.635 min Scan# 603


Delta R.T. 0.006 min

Lab File: VX021184.D

Acq: 12 Mar 2021 16:33

Tgt Ion:	43	Resp:	716666
Ion Ratio	Lower	Upper	
43	100		
57	7.9	7.2	10.8
72	23.9	23.7	35.5





#38

Methylcyclohexane

Concen: 0.70 ug/l

RT: 7.870 min Scan# 1153

Delta R.T. 0.430 min

Lab File: VX021184.D

Acq: 12 Mar 2021 16:33

Tgt Ion: 83 Resp: 1106

Ion Ratio Lower Upper

83 100

55 0.0 37.4 112.1#





98 0.0 24.7 74.1#

Instrument :

MSVOA_X

ClientSampleId :

D2470-D2800(WATER)



#40

Dibromomethane

Concen: 0.22 ug/l

RT: 7.635 min Scan# 1113

Delta R.T. 0.000 min

Lab File: VX021184.D

Acq: 12 Mar 2021 16:33




Tgt Ion: 93 Resp: 168


Ion Ratio Lower Upper

93 100

95 93.5 0.0 167.4

174 71.4 0.0 211.8









#41
Bromodichloromethane
Concen: 0.68 ug/l
RT: 7.870 min Scan# 1153
Delta R.T. -0.006 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion: 83 Resp: 1106
Ion Ratio Lower Upper
83 100
85 67.2 51.4 77.0
127 0.0 7.1 10.7#



#51
Ethyl methacrylate
Concen: 0.27 ug/l
RT: 8.700 min Scan# 1294
Delta R.T. -0.459 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Tgt Ion: 69 Resp: 444
Ion Ratio Lower Upper
69 100
41 96.2 29.1 87.4#
39 37.2 15.4 46.1









#56
Bromoform
Concen: 0.21 ug/l
RT: 11.124 min Scan# 1706
Delta R.T. 0.283 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion:173 Resp: 177
Ion Ratio Lower Upper
173 100
175 1138.4 38.1 57.1#
254 0.0 8.2 12.4#



#57
Chlorobenzene-d5
Concen: 30.00 ug/l
RT: 10.094 min Scan# 1531
Delta R.T. 0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Tgt Ion:117 Resp: 73535
Ion Ratio Lower Upper
117 100
82 60.1 45.0 67.6
119 31.8 25.6 38.4









#58
4-Methyl-2-Pentanone
Concen: 0.54 ug/l
RT: 8.618 min Scan# 1280
Delta R.T. 0.001 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion: 43 Resp: 798
Ion Ratio Lower Upper
43 100
58 34.0 35.2 52.8#
85 8.1 16.8 25.2#



#59
2-Hexanone
Concen: 0.24 ug/l
RT: 9.471 min Scan# 1425
Delta R.T. 0.001 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Tgt Ion: 43 Resp: 265
Ion Ratio Lower Upper
43 100
58 16.2 49.0 73.6#
57 0.0 17.1 25.7#









#60
4-Bromofluorobenzene
Concen: 28.25 ug/l
RT: 11.118 min Scan# 1705
Delta R.T. -0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion: 95 Resp: 37976
Ion Ratio Lower Upper
95 100
174 70.9 63.0 94.4
176 68.5 59.9 89.9



#61
Tetrachloroethene
Concen: 0.20 ug/l
RT: 9.324 min Scan# 1400
Delta R.T. 0.006 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Tgt Ion: 164 Resp: 195
Ion Ratio Lower Upper
164 100
166 70.2 102.8 154.2#
129 54.8 75.8 113.8#
131 93.5 77.8 116.6









#62
Toluene
Concen: 1.49 ug/l
RT: 8.765 min Scan# 1305
Delta R.T. 0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion: 91 Resp: 6066
Ion Ratio Lower Upper
91 100
92 55.2 47.2 70.8



#63
Toluene-d8
Concen: 30.83 ug/l
RT: 8.694 min Scan# 1293
Delta R.T. -0.006 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Tgt Ion: 98 Resp: 102854
Ion Ratio Lower Upper
98 100
100 64.4 53.0 79.4









#66
Ethyl Benzene
Concen: 48.77 ug/l
RT: 10.235 min Scan# 1555
Delta R.T. 0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion: 91 Resp: 228641
Ion Ratio Lower Upper
91 100
106 30.2 25.0 37.4



#67
m/p-Xylenes
Concen: 205.11 ug/l
RT: 10.341 min Scan# 1573
Delta R.T. 0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Tgt Ion:106 Resp: 346147
Ion Ratio Lower Upper
106 100
91 217.5 164.0 246.0









#68
o-Xylene
Concen: 119.72 ug/l
RT: 10.683 min Scan# 1631
Delta R.T. 0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion:106 Resp: 195600
Ion Ratio Lower Upper
106 100
91 227.0 106.5 319.6



#69
Styrene
Concen: 3.70 ug/l
RT: 10.683 min Scan# 1631
Delta R.T. -0.012 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Tgt Ion:104 Resp: 10264
Ion Ratio Lower Upper
104 100
78 335.5 39.2 58.8#
103 263.8 43.0 64.4#





#70

Isopropylbenzene

Concen: 0.45 ug/l

RT: 11.000 min Scan# 1685

Delta R.T. 0.000 min

Lab File: VX021184.D

Acq: 12 Mar 2021 16:33

Tgt Ion:105 Resp: 2012

Ion Ratio Lower Upper

105 100





120 26.9 18.4 34.2

Instrument :

MSVOA_X

ClientSampleId :

D2470-D2800(WATER)



#77

t-1,4-Dichloro-2-butene

Concen: 44.82 ug/l

RT: 11.118 min Scan# 1705

Delta R.T. 0.059 min

Lab File: VX021184.D

Acq: 12 Mar 2021 16:33




Tgt Ion: 75 Resp: 21229


Ion Ratio Lower Upper

75 100

53 0.0 48.4 145.0#

89 0.0 22.7 68.0#









#79
tert-butylbenzene
Concen: 0.29 ug/l
RT: 12.047 min Scan# 1863
Delta R.T. 0.294 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion:119	Resp:	1031
Ion Ratio	Lower	Upper
119	100	
91	30.6	0.0 118.0
134	23.1	0.0 48.0



#80
1,2,4-Trimethylbenzene
Concen: 0.28 ug/l
RT: 11.794 min Scan# 1820
Delta R.T. 0.006 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Tgt Ion:105	Resp:	1060
Ion Ratio	Lower	Upper
105	100	
120	33.8	0.0 89.8









#82
p-Isopropyltoluene
Concen: 0.27 ug/l
RT: 12.047 min Scan# 1863
Delta R.T. 0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion:119	Resp:	1031
Ion	Ratio	Lower Upper
119	100	
134	23.1	0.0 52.8
91	30.6	0.0 47.4



#83
1,3-Dichlorobenzene
Concen: 0.32 ug/l
RT: 12.006 min Scan# 1856
Delta R.T. 0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33

Tgt Ion:146	Resp:	606
Ion	Ratio	Lower Upper
146	100	
111	42.2	20.3 60.9
148	55.0	31.9 95.9









#84
1,4-Dichlorobenzene
Concen: 0.43 ug/l
RT: 12.083 min Scan# 1869
Delta R.T. 0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)


Tgt Ion:146 Resp: 827
Ion Ratio Lower Upper
146 100
111 37.8 19.1 57.3
148 56.2 31.9 95.5



#87
1,2-Dichlorobenzene
Concen: 0.45 ug/l
RT: 12.083 min Scan# 1869
Delta R.T. -0.294 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33


Tgt Ion:146 Resp: 827
Ion Ratio Lower Upper
146 100
111 37.8 20.8 62.5
148 56.2 32.4 97.0








#89
1,2,4-Trichlorobenzene
Concen: 0.50 ug/l
RT: 13.624 min Scan# 2131
Delta R.T. -0.006 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33


Instrument :
MSVOA_X
ClientSampleId :
D2470-D2800(WATER)





Tgt Ion:180	Resp:	590
Ion	Ratio	Lower Upper
180	100	
182	84.4	77.9 116.9
145	31.9	24.3 36.5




#91
Naphthalene
Concen: 1.78 ug/l
RT: 13.818 min Scan# 2164
Delta R.T. 0.000 min
Lab File: VX021184.D
Acq: 12 Mar 2021 16:33




Tgt Ion:128	Resp:	6719
Ion	Ratio	Lower Upper
128	100	
127	13.3	10.2 15.4
129	10.8	8.8 13.2





#92
 1,2,3-Trichlorobenzene
 Concen: 0.52 ug/l
 RT: 13.624 min Scan# 2131
 Delta R.T. -0.376 min
 Lab File: VX021184.D
 Acq: 12 Mar 2021 16:33

Instrument :
 MSVOA_X
 ClientSampleId :
 D2470-D2800(WATER)



Tgt Ion:	180	Resp:	590
Ion	Ratio	Lower	Upper
180	100		
182	84.4	0.0	192.4
145	31.9	0.0	66.4

