

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031221\
 Data File : VX021180.D
 Acq On : 12 Mar 2021 13:59
 Operator : JC/MD
 Sample : VX0312WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0312WBL01

Quant Time: Mar 13 01:05:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X031221W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Mar 12 12:55:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.976	128	15666	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.829	114	86640	30.00	ug/l	# 0.00
57) Chlorobenzene-d5	10.094	117	72944	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.029	65	44132	30.61	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 102.03%	
60) 4-Bromofluorobenzene	11.118	95	36069	27.04	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 90.13%	
63) Toluene-d8	8.694	98	104270	31.51	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 105.03%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031221\
Data File : VX021180.D
Acq On : 12 Mar 2021 13:59
Operator : JC/MD
Sample : VX0312WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0312WBL01

Quant Time: Mar 13 01:05:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X031221W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Mar 12 12:55:27 2021
Response via : Initial Calibration

