



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031221\
 Data File : VX021171.D
 Acq On : 12 Mar 2021 09:49
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X031221W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Fri Mar 12 12:55:27 2021



AutoFind: Scans 1705, 1706, 1707; Background Corrected with Scan 1699

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.1	4501	PASS
75	95	30	60	57.2	11659	PASS
95	95	100	100	100.0	20375	PASS
96	95	5	9	6.9	1399	PASS
173	174	0.00	2	1.0	146	PASS
174	95	50	100	70.4	14354	PASS
175	174	5	9	7.9	1134	PASS
176	174	95	101	98.8	14187	PASS
177	176	5	9	6.5	925	PASS