

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031225\
 Data File : VX045243.D
 Acq On : 12 Mar 2025 13:01
 Operator : JC/MD
 Sample : Q1531-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP01-20250306

Quant Time: Mar 13 01:40:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	68183	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135625	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	124508	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51386	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58347	53.798	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.600%		
35) Dibromofluoromethane	5.385	113	47002	51.828	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.660%		
50) Toluene-d8	8.647	98	167501	50.946	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.900%		
62) 4-Bromofluorobenzene	11.079	95	57978	53.216	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 106.440%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.318	101	2280	2.877	ug/l	98
12) 1,1-Dichloroethene	2.306	96	589	0.703	ug/l #	66
16) Acetone	2.386	43	878	1.745	ug/l	97
21) trans-1,2-Dichloroethene	3.087	96	544	0.657	ug/l #	66
27) cis-1,2-Dichloroethene	4.495	96	2228	2.181	ug/l	96
44) Trichloroethene	7.122	130	340297	369.016	ug/l	100
64) Tetrachloroethene	9.275	164	1934	2.425	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031225\
Data File : VX045243.D
Acq On : 12 Mar 2025 13:01
Operator : JC/MD
Sample : Q1531-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP01-20250306

Quant Time: Mar 13 01:40:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

