

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031225\
 Data File : VX045247.D
 Acq On : 12 Mar 2025 14:35
 Operator : JC/MD
 Sample : Q1531-19DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D2-20250307DL

Quant Time: Mar 13 01:41:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	64391	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128861	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50133	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57148	55.796	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.600%
35) Dibromofluoromethane	5.385	113	44957	52.175	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.340%
50) Toluene-d8	8.647	98	163636	52.383	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.760%
62) 4-Bromofluorobenzene	11.079	95	56446	54.530	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	109.060%
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.495	96	4459	4.621	ug/l	94
44) Trichloroethene	7.123	130	35712	40.759	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031225\
Data File : VX045247.D
Acq On : 12 Mar 2025 14:35
Operator : JC/MD
Sample : Q1531-19DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D2-20250307DL

Quant Time: Mar 13 01:41:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

