

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031225\
 Data File : VX045251.D
 Acq On : 12 Mar 2025 16:08
 Operator : JC/MD
 Sample : Q1539-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TAP-TB-04-031025-T2

Quant Time: Mar 13 01:43:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	64979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127269	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46701	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56511	54.675	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.340%	
35) Dibromofluoromethane	5.379	113	44507	52.299	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.600%	
50) Toluene-d8	8.647	98	161479	52.339	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.680%	
62) 4-Bromofluorobenzene	11.079	95	54028	52.847	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.700%	
Target Compounds						
16) Acetone	2.386	43	1340	2.794	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031225\
Data File : VX045251.D
Acq On : 12 Mar 2025 16:08
Operator : JC/MD
Sample : Q1539-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TAP-TB-04-031025-T2

Quant Time: Mar 13 01:43:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

