

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031319\
 Data File : VX008042.D
 Acq On : 13 Mar 2019 12:12
 Operator : JC/SP
 Sample : K1840-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-685-687

Quant Time: Mar 13 13:00:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251397	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	393979	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	375510	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184862	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	133790	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
35) Dibromofluoromethane	5.50	113	126117	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
50) Toluene-d8	8.71	98	474818	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
62) 4-Bromofluorobenzene	11.13	95	179268	55.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.48%	
Target Compounds						
16) Acetone	2.45	43	5094	3.319	ug/l #	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031319\
Data File : VX008042.D
Acq On : 13 Mar 2019 12:12
Operator : JC/SP
Sample : K1840-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-685-687

Quant Time: Mar 13 13:00:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

