

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031319\
 Data File : VX008044.D
 Acq On : 13 Mar 2019 13:05
 Operator : JC/SP
 Sample : K1840-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-700-702

Quant Time: Mar 13 13:27:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254925	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	398478	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	378494	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186655	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	136363	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
35) Dibromofluoromethane	5.51	113	126711	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
50) Toluene-d8	8.71	98	481398	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
62) 4-Bromofluorobenzene	11.13	95	180578	55.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.04%	
Target Compounds						
16) Acetone	2.45	43	8509	5.467	ug/l	99
25) 2-Butanone	4.70	43	4519	2.051	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031319\
Data File : VX008044.D
Acq On : 13 Mar 2019 13:05
Operator : JC/SP
Sample : K1840-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-700-702

Quant Time: Mar 13 13:27:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

