

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031319\
 Data File : VX008045.D
 Acq On : 13 Mar 2019 13:31
 Operator : JC/SP
 Sample : K1840-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-660-662

Quant Time: Mar 13 13:52:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	255770	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	397768	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	381648	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	185837	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	135465	50.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.42%	
35) Dibromofluoromethane	5.51	113	126199	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
50) Toluene-d8	8.71	98	485858	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
62) 4-Bromofluorobenzene	11.13	95	182001	55.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.10%	
Target Compounds						
16) Acetone	2.45	43	7212	4.618	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031319\
Data File : VX008045.D
Acq On : 13 Mar 2019 13:31
Operator : JC/SP
Sample : K1840-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-660-662

Quant Time: Mar 13 13:52:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

