

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031319\
 Data File : VX008050.D
 Acq On : 13 Mar 2019 15:44
 Operator : JC/SP
 Sample : K1829-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CPA-TB1-M20W2-030419

Quant Time: Mar 14 10:23:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	253339	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	391927	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	373064	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	179619	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	134318	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
35) Dibromofluoromethane	5.51	113	126359	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
50) Toluene-d8	8.71	98	477996	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
62) 4-Bromofluorobenzene	11.13	95	176282	54.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.22%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031319\
Data File : VX008050.D
Acq On : 13 Mar 2019 15:44
Operator : JC/SP
Sample : K1829-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CPA-TB1-M20W2-030419

Quant Time: Mar 14 10:23:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

