

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031319\
 Data File : VX008057.D
 Acq On : 13 Mar 2019 18:49
 Operator : JC/SP
 Sample : K1828-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CPA-CPMW10D-M20W2-030519

Quant Time: Mar 14 04:52:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250217	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	388604	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	370466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	179020	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	133677	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
35) Dibromofluoromethane	5.50	113	125361	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
50) Toluene-d8	8.71	98	472444	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
62) 4-Bromofluorobenzene	11.13	95	173013	54.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.10%	
Target Compounds						
16) Acetone	2.45	43	3520	2.304	ug/l	96
64) Tetrachloroethene	9.34	164	5349	1.742	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031319\
Data File : VX008057.D
Acq On : 13 Mar 2019 18:49
Operator : JC/SP
Sample : K1828-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CPA-CPMW10D-M20W2-030519

Quant Time: Mar 14 04:52:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

