

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031320\
 Data File : VX015248.D
 Acq On : 13 Mar 2020 10:49
 Operator : JC/SP
 Sample : L1903-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20200307

Quant Time: Mar 16 08:48:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	290631	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	455962	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	438275	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	219648	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	164549	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
35) Dibromofluoromethane	5.50	113	141684	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
50) Toluene-d8	8.71	98	555246	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
62) 4-Bromofluorobenzene	11.13	95	200942	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
Target Compounds						
20) Methylene Chloride	2.85	84	7206	2.205	ug/l #	84
52) Toluene	8.78	92	15360	2.009	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031320\
Data File : VX015248.D
Acq On : 13 Mar 2020 10:49
Operator : JC/SP
Sample : L1903-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20200307

Quant Time: Mar 16 08:48:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

