

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031320\
 Data File : VX015261.D
 Acq On : 13 Mar 2020 15:43
 Operator : JC/SP
 Sample : L1903-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D2-20200309

Quant Time: Mar 16 09:23:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	251654	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	397210	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	383983	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	198494	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	142676	47.79	ug/l	0.00
Spiked Amount			Recovery	=	95.58%	
35) Dibromofluoromethane	5.49	113	120805	48.21	ug/l	0.00
Spiked Amount			Recovery	=	96.42%	
50) Toluene-d8	8.71	98	487013	51.09	ug/l	0.00
Spiked Amount			Recovery	=	102.18%	
62) 4-Bromofluorobenzene	11.13	95	180234	53.01	ug/l	0.00
Spiked Amount			Recovery	=	106.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.38	101	50596	21.352	ug/l	98
12) 1,1-Dichloroethene	2.37	96	6445	2.635	ug/l	94
27) cis-1,2-Dichloroethene	4.59	96	9504	3.244	ug/l	86
44) Trichloroethene	7.21	130	2058874	678.593	ug/l	98
49) 1,4-Dioxane	7.75	88	1607	24.371	ug/l	94
55) 1,1,2-Trichloroethane	9.21	97	1650	0.602	ug/l	91
64) Tetrachloroethene	9.33	164	16744	5.274	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031320\
Data File : VX015261.D
Acq On : 13 Mar 2020 15:43
Operator : JC/SP
Sample : L1903-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D2-20200309

Quant Time: Mar 16 09:23:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

