


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX031320\
 Data File : VX015243.D
 Acq On : 13 Mar 2020 08:21
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Title : SW846 8260
 Last Update : Wed Mar 04 14:19:10 2020



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1671

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	18514	PASS
75	95	30	60	47.6	46101	PASS
95	95	100	100	100.0	96923	PASS
96	95	5	9	6.6	6411	PASS
173	174	0.00	2	1.5	1268	PASS
174	95	50	100	86.8	84104	PASS
175	174	5	9	7.8	6594	PASS
176	174	95	101	97.9	82325	PASS
177	176	5	9	6.3	5187	PASS