

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031323\
 Data File : VX034427.D
 Acq On : 13 Mar 2023 13:03
 Operator : JC/MD
 Sample : 01879-10 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP03-20230308

Quant Time: Mar 14 02:23:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Mar 13 10:29:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	278139	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	470150	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	408479	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171811	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	181057	44.746	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	89.500%	
35) Dibromofluoromethane	5.385	113	150019	48.328	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.660%	
50) Toluene-d8	8.647	98	567528	48.975	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	97.960%	
62) 4-Bromofluorobenzene	11.079	95	189801	40.550	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	81.100%#	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1237	0.400	ug/l	93
12) 1,1-Dichloroethene	2.325	96	1276	0.419	ug/l	89
17) Carbon Disulfide	2.514	76	2670	0.312	ug/l	99
20) Methylene Chloride	2.788	84	1405	0.380	ug/l #	86
44) Trichloroethene	7.123	130	636859	175.737	ug/l	99
64) Tetrachloroethene	9.281	164	807	0.294	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031323\
Data File : VX034427.D
Acq On : 13 Mar 2023 13:03
Operator : JC/MD
Sample : 01879-10 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP03-20230308

Quant Time: Mar 14 02:23:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X031023W.M
Quant Title : SW846 8260
QLast Update : Mon Mar 13 10:29:32 2023
Response via : Initial Calibration

