

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX031323\  
 Data File : VX034429.D  
 Acq On : 13 Mar 2023 13:50  
 Operator : JC/MD  
 Sample : 01879-16 20X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 DUP01-20230308

Quant Time: Mar 14 02:24:33 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031023W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Mar 13 10:29:32 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 267516   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 464951   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 412990   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 188794   | 50.000   | ug/l   | 0.00     |
|                             |        |       |          |          |        |          |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 178003   | 45.738   | ug/l   | 0.01     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =      | 91.480%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 145018   | 47.240   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 94.480%  |
| 50) Toluene-d8              | 8.647  | 98    | 560916   | 48.946   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =      | 97.900%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 200776   | 43.374   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =      | 86.740%  |
|                             |        |       |          |          |        |          |
| Target Compounds            |        |       |          |          |        |          |
| 20) Methylene Chloride      | 2.794  | 84    | 1190     | 0.334    | ug/l # | 81       |
| 44) Trichloroethene         | 7.123  | 130   | 4291     | 1.197    | ug/l   | 99       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX031323\  
Data File : VX034429.D  
Acq On : 13 Mar 2023 13:50  
Operator : JC/MD  
Sample : 01879-16 20X  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
DUP01-20230308

Quant Time: Mar 14 02:24:33 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X031023W.M  
Quant Title : SW846 8260  
QLast Update : Mon Mar 13 10:29:32 2023  
Response via : Initial Calibration

