

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
 Data File : VX045263.D
 Acq On : 13 Mar 2025 12:50
 Operator : JC/MD
 Sample : Q1526-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Mar 14 12:27:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	70415	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48467	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60076	53.637	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.280%
35) Dibromofluoromethane	5.379	113	47758	52.578	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.160%
50) Toluene-d8	8.647	98	170597	51.806	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.620%
62) 4-Bromofluorobenzene	11.079	95	57062	52.293	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
Data File : VX045263.D
Acq On : 13 Mar 2025 12:50
Operator : JC/MD
Sample : Q1526-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SOIL-REF-6

Quant Time: Mar 14 12:27:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

