

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
 Data File : VX045268.D
 Acq On : 13 Mar 2025 14:47
 Operator : JC/MD
 Sample : Q1559-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20250310

Quant Time: Mar 14 12:28:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	68915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	136302	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	126125	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51157	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59825	54.575	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.160%	
35) Dibromofluoromethane	5.385	113	48791	53.533	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.060%	
50) Toluene-d8	8.647	98	169613	51.333	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.660%	
62) 4-Bromofluorobenzene	11.079	95	58569	53.492	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.980%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1959	2.258	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.325	101	26617	33.231	ug/l	98
12) 1,1-Dichloroethene	2.313	96	10780	12.730	ug/l	96
16) Acetone	2.380	43	1087	2.137	ug/l	92
24) 1,1-Dichloroethane	3.611	63	2848	1.658	ug/l #	95
27) cis-1,2-Dichloroethene	4.489	96	3543	3.431	ug/l	93
30) Chloroform	5.093	83	3627	2.125	ug/l	98
38) Carbon Tetrachloride	5.660	117	4609	3.632	ug/l	86
44) Trichloroethene	7.123	130	502161	541.835	ug/l	99
55) 1,1,2-Trichloroethane	9.159	97	1119	1.174	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
Data File : VX045268.D
Acq On : 13 Mar 2025 14:47
Operator : JC/MD
Sample : Q1559-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20250310

Quant Time: Mar 14 12:28:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

