

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
 Data File : VX045272.D
 Acq On : 13 Mar 2025 16:19
 Operator : JC/MD
 Sample : Q1559-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT149S1-20250310

Quant Time: Mar 14 12:29:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129573	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121218	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	57875	55.159	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.320%	
35) Dibromofluoromethane	5.385	113	45635	52.671	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.340%	
50) Toluene-d8	8.647	98	164879	52.491	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.980%	
62) 4-Bromofluorobenzene	11.079	95	57650	55.387	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 110.780%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.319	101	1585	2.067	ug/l	96
16) Acetone	2.380	43	1324	2.720	ug/l	93
19) Methyl tert-butyl Ether	3.111	73	29429	10.822	ug/l	98
44) Trichloroethene	7.129	130	4675	5.306	ug/l	99
64) Tetrachloroethene	9.275	164	1214	1.564	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
Data File : VX045272.D
Acq On : 13 Mar 2025 16:19
Operator : JC/MD
Sample : Q1559-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT149S1-20250310

Quant Time: Mar 14 12:29:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

