

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
 Data File : VX045273.D
 Acq On : 13 Mar 2025 16:43
 Operator : JC/MD
 Sample : Q1559-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D3-20250310

Quant Time: Mar 14 12:29:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	69221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	137405	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	125442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54042	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60400	54.856	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.720%	
35) Dibromofluoromethane	5.379	113	46739	50.870	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.740%	
50) Toluene-d8	8.647	98	168759	50.664	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.320%	
62) 4-Bromofluorobenzene	11.079	95	58286	52.806	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.620%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.319	101	146958	182.665	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	1527	1.472	ug/l	89
44) Trichloroethene	7.123	130	79338	84.919	ug/l	99
64) Tetrachloroethene	9.275	164	9347	11.634	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
Data File : VX045273.D
Acq On : 13 Mar 2025 16:43
Operator : JC/MD
Sample : Q1559-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D3-20250310

Quant Time: Mar 14 12:29:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

