

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
 Data File : VX045275.D
 Acq On : 13 Mar 2025 17:29
 Operator : JC/MD
 Sample : Q1559-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW10-MW01S-20250311

Quant Time: Mar 14 12:29:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	68205	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	136636	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	127965	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60861	56.098	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.200%
35) Dibromofluoromethane	5.379	113	47581	52.078	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.160%
50) Toluene-d8	8.647	98	171109	51.659	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.320%
62) 4-Bromofluorobenzene	11.079	95	59369	54.090	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.180%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1464	2.908	ug/l	89
19) Methyl tert-butyl Ether	3.117	73	3170	1.127	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	858	0.839	ug/l	87
44) Trichloroethene	7.129	130	882	0.949	ug/l	71
64) Tetrachloroethene	9.275	164	4731	5.773	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
Data File : VX045275.D
Acq On : 13 Mar 2025 17:29
Operator : JC/MD
Sample : Q1559-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW10-MW01S-20250311

Quant Time: Mar 14 12:29:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

