

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
 Data File : VX045276.D
 Acq On : 13 Mar 2025 17:52
 Operator : JC/MD
 Sample : Q1559-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT158S1-20250311

Quant Time: Mar 14 12:29:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	70011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	136273	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50668	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60857	54.648	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.300%			
35) Dibromofluoromethane	5.379	113	48070	52.753	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.500%			
50) Toluene-d8	8.647	98	171285	51.850	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.700%			
62) 4-Bromofluorobenzene	11.079	95	57293	52.338	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 104.680%			
Target Compounds						
16) Acetone	2.386	43	876	1.695	ug/l	95
24) 1,1-Dichloroethane	3.605	63	2660	1.524	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
Data File : VX045276.D
Acq On : 13 Mar 2025 17:52
Operator : JC/MD
Sample : Q1559-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT158S1-20250311

Quant Time: Mar 14 12:29:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

