

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
 Data File : VX045281.D
 Acq On : 13 Mar 2025 19:48
 Operator : JC/MD
 Sample : Q1559-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D1-20250312

Quant Time: Mar 15 05:39:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71280	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	137818	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51927	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60978	53.781	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.560%	
35) Dibromofluoromethane	5.385	113	48880	53.041	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.080%	
50) Toluene-d8	8.647	98	172478	51.625	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.260%	
62) 4-Bromofluorobenzene	11.079	95	58312	52.671	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.340%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	1038	1.253	ug/l	98
44) Trichloroethene	7.123	130	40894	43.640	ug/l	99
64) Tetrachloroethene	9.275	164	2218	2.755	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
Data File : VX045281.D
Acq On : 13 Mar 2025 19:48
Operator : JC/MD
Sample : Q1559-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D1-20250312

Quant Time: Mar 15 05:39:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

