

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
 Data File : VX045282.D
 Acq On : 13 Mar 2025 20:12
 Operator : JC/MD
 Sample : Q1559-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20250312

Quant Time: Mar 14 12:30:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	64110	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125932	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49199	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56440	55.346	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.700%	
35) Dibromofluoromethane	5.379	113	44382	52.706	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.420%	
50) Toluene-d8	8.647	98	156681	51.323	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.640%	
62) 4-Bromofluorobenzene	11.079	95	54299	53.676	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.360%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.318	101	1054	1.415	ug/l	95
16) Acetone	2.386	43	1457	3.079	ug/l	91
44) Trichloroethene	7.129	130	10603	12.383	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031325\
Data File : VX045282.D
Acq On : 13 Mar 2025 20:12
Operator : JC/MD
Sample : Q1559-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D1-20250312

Quant Time: Mar 14 12:30:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

