

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031419\
 Data File : VX008073.D
 Acq On : 14 Mar 2019 16:19
 Operator : JC/SP
 Sample : K1847-07
 Misc : 5.90g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUP-7B-G-7-031119

Quant Time: Mar 15 04:43:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	260440	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	399849	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	395052	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	218991	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	130451	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
35) Dibromofluoromethane	5.50	113	120007	47.33	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.66%	
50) Toluene-d8	8.71	98	487649	50.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.14%	
62) 4-Bromofluorobenzene	11.14	95	198417	57.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.86%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031419\
Data File : VX008073.D
Acq On : 14 Mar 2019 16:19
Operator : JC/SP
Sample : K1847-07
Misc : 5.90g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SUP-7B-G-7-031119

Quant Time: Mar 15 04:43:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

