

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031419\
 Data File : VX008086.D
 Acq On : 14 Mar 2019 22:53
 Operator : JC/SP
 Sample : K1829-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CPA-CPMW10S-M20W2-030519

Quant Time: Mar 15 05:19:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	312946	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	483000	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	454910	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	225735	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	157477	49.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.76%	
35) Dibromofluoromethane	5.51	113	152772	49.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.76%	
50) Toluene-d8	8.71	98	577127	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	11.13	95	214476	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
Target Compounds						
16) Acetone	2.45	43	5898	3.957	ug/l	98
27) cis-1,2-Dichloroethene	4.61	96	1443	0.472	ug/l	99
30) Chloroform	5.21	83	1840	0.384	ug/l	99
44) Trichloroethene	7.21	130	2693	0.767	ug/l	94
64) Tetrachloroethene	9.34	164	60709	16.471	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031419\
Data File : VX008086.D
Acq On : 14 Mar 2019 22:53
Operator : JC/SP
Sample : K1829-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CPA-CPMW10S-M20W2-030519

Quant Time: Mar 15 05:19:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

