

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031419\
 Data File : VX008088.D
 Acq On : 14 Mar 2019 23:46
 Operator : JC/SP
 Sample : K1839-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE109D1-20190307

Quant Time: Mar 15 05:24:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	315275	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	476220	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	457120	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	223188	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	155529	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
35) Dibromofluoromethane	5.51	113	150111	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
50) Toluene-d8	8.71	98	577411	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
62) 4-Bromofluorobenzene	11.13	95	213883	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	1414	0.556	ug/l	93
16) Acetone	2.45	43	3729	2.484	ug/l #	87
44) Trichloroethene	7.21	130	82517	23.823	ug/l	96
64) Tetrachloroethene	9.34	164	2438	0.658	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031419\
Data File : VX008088.D
Acq On : 14 Mar 2019 23:46
Operator : JC/SP
Sample : K1839-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE109D1-20190307

Quant Time: Mar 15 05:24:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

