

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031419\
 Data File : VX008090.D
 Acq On : 15 Mar 2019 00:40
 Operator : JC/SP
 Sample : K1839-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D2-20190307

Quant Time: Mar 15 05:28:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	299866	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	458553	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	436007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	214234	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	152432	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
35) Dibromofluoromethane	5.51	113	146503	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
50) Toluene-d8	8.71	98	549986	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
62) 4-Bromofluorobenzene	11.13	95	201483	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
Target Compounds						
16) Acetone	2.45	43	4450	3.116	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	27904	9.529	ug/l	86
30) Chloroform	5.22	83	4940	1.075	ug/l	97
44) Trichloroethene	7.21	130	127933	38.357	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031419\
Data File : VX008090.D
Acq On : 15 Mar 2019 00:40
Operator : JC/SP
Sample : K1839-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D2-20190307

Quant Time: Mar 15 05:28:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

