

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031419\
 Data File : VX008097.D
 Acq On : 15 Mar 2019 03:46
 Operator : JC/SP
 Sample : K1839-04DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT101D2-20190307DL

Quant Time: Mar 15 05:43:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 03:56:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	254352	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	388321	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	368097	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	195438	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	129257	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
35) Dibromofluoromethane	5.50	113	124493	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
50) Toluene-d8	8.71	98	468297	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
62) 4-Bromofluorobenzene	11.13	95	175675	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	4297	2.094	ug/l	98
12) 1,1-Dichloroethene	2.37	96	1304	0.672	ug/l	84
44) Trichloroethene	7.21	130	280883	99.447	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031419\
Data File : VX008097.D
Acq On : 15 Mar 2019 03:46
Operator : JC/SP
Sample : K1839-04DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT101D2-20190307DL

Quant Time: Mar 15 05:43:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X031419W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 03:56:35 2019
Response via : Initial Calibration

